

Allweighs Weighing Services  
Westend Grange  
West End Road  
Epworth  
Doncaster  
DN9 1LA  
Tel: 01427 875 867  
Web: www.allweighs.co.uk



Certificate No: KCS:

202401090

Location:

Weighbridge

Manufacturer:

Avery / Dini Argeo 3590 ET

Serial/Plant No:

83626665

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

24 October 2024

Next due:

October 2025

## Calibration Certificate of Accuracy

Business Name:

AMA Storage

Address:

Rawmarsh Road  
Rotherham  
South Yorkshire

Contact Name:

### Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6700	0			
2	6700	0			
3	6700	0			
4	6700	0			

☒ Four loadcell

☐ Six loadcell

☐ Single loadcell

☐ Eight loadcell

### As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23680	23680	0
33680	33680	0
44400	44400	0


Repeatability

44400	44400	0
-------	-------	---

Notes:

FOUND TO BE CORRECT

Engineer  
Name:



### Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23680	23680	0
33680	33680	0
44400	44400	0

Repeatability

44400	44400	0
-------	-------	---

Traceable Weights Used:

W24017

Trading  
Standards  
Maximum  
Tolerance:

60

+/-

60

kg

Calibration method and equipment used  
traceable to national standards.

QMF17 Issue 5 14/07/2020